



INTERNATIONAL STANDARD ISO 11196:1995 TECHNICAL CORRIGENDUM 1

Published 1997-02-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Anaesthetic gas monitors

TECHNICAL CORRIGENDUM 1

Dispositifs de contrôle de gaz d'anesthésie

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to International Standard ISO 11196:1995 was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 1, *Breathing attachments and anaesthetic machines*.

Page 18

Replace existing table 2 with new table 2 below, in which the value "5" has been added in column 2, row 3; the value "5" has been replaced with the value "15" in column 8, row 16; and in table footnote 1) the value "5 %" for desflurane has been replaced with the value "15 %".

ICS 11.040.10

Ref. No. ISO 11196:1995/Cor.1:1997(E)

Descriptors: medical equipment, anaesthetic equipment, medical gases, gas analyzers, classification, specifications, safety requirements, tests, identification methods, marking, instructions for use.

© ISO 1997

Printed in Switzerland